

Comparison between Conventional Radiography (IOPA) and Digital Radiography using Bitewing Technique in Detecting the Depth of Alveolar Bone Loss

Satvinder Singh¹

¹ Indira Gandhi Dental College Jammu

Received: 11 April 2015 Accepted: 30 April 2015 Published: 15 May 2015

Abstract

Objective: To assess the depth of alveolar bone loss by using Conventional radiography (IOPA) and Digital radiography (RVG) technique in periodontitis as it affects the connective tissue attachment and supporting bone around the teeth. **Methods:** The study was carried out on 40 males and 10 females aged between 20 - 65 years who have generalized mild to severe chronic periodontitis. A series of conventional bitewing radiographs and digital bitewing radiographs (15,16,17,25,26,27,35,36,37,45,46,47) were taken for each patient. The Statistical software namely SPSS version 16.0 was used for data analysis. Paired t-test was performed on all the variables to evaluate between both the groups at $p < 0.05$. **Results:** The overall results showed the mean statistical difference between both the conventional and digital bitewing radiographs as 0.4595. It was observed that overall digital bitewing radiographs averaged about 0.4mm greater bone loss than conventional bitewing radiographs. **Conclusion:** It was concluded that digital radiographs showed better results when compared to conventional radiographs in terms of alveolar bone loss as RVG has superior image recording capabilities. **Methods:**The study was carried out on 40 males and 10 females aged between 20 -65 years who have generalized mild to severe chronic periodontitis. A series of conventional bitewing radiographs and digital bitewing radiographs (15,16,17,25,26,27,35,36,37,45,46,47) were taken for each patient. The Statistical software namely SPSS version 16.0 was used for data analysis. Paired t-test was performed on all the variables to evaluate between both the groups at $p < 0.05$. **Results:**The overall results showed the mean statistical difference between both the conventional and digital bitewing radiographs as 0.4595. It was observed that overall digital bitewing radiographs averaged about 0.4mm greater bone loss than conventional bitewing radiographs. **Conclusion:**It was concluded that digital radiographs showed better results when compared to conventional radiographs in terms of alveolar bone loss as RVG has superior image recording capabilities.

Index terms—

1 Introduction

Diagnosis is the process of assessing a patient's health as well as ensuing opinions formulated by clinicians. Oral diagnosis is the art of using scientific knowledge to determine the nature of oral diseases and distinguishing it from other diseases [1].

Radiography is a well established procedure in daily dental practice and is still the most basic and an important diagnostic tool available. Radiographs play an integral role in the assessment of periodontal diseases.

5 EXCLUSION CRITERIA

41 Conventional bitewing and intra oral periapical radiographs are commonly used to detect alveolar bone loss
42 associated with periodontal disease. They provide unique information about the status of the periodontium and
43 a permanent record of the bone throughout the course of the disease. However, the quality of an X-ray sensitive
44 film can be affected by multiple variables such as improper exposure, under-or overdeveloping and poor fixing
45 [2].

46 Over the past few years, systems that can generate radiographic digital images without the need for radiography
47 film have become available for use in clinical practice and are gaining popularity among practitioners. Such digital
48 radiography can also reduce the radiation exposure. One of the most useful advantages of digital radiography is
49 the knack it provides to the clinicians to send images to practitioners in a matter of minutes, for which it has
50 become widely accepted as an alternative to film-based radiography [1].

51 A few studies (most of which were performed in vitro) that examined the use of digital radiography in
52 evaluating crestal alveolar bone loss associated with periodontal disease listed out few of its main advantages over
53 conventional radiography to be speed, convenience, lack of dark room procedures, image improvement tools and
54 dose reduction. Thus the implementation of digital radiography in a dental practice seems to provide a solution
55 for the future imaging requirements [2].

56 Nevertheless, according to some authors intra oral direct digital radiographs are not an equivalent substitute
57 for conventional radiographs in evaluating alveolar bone levels [3]. Few of its disadvantages include cost of the
58 device, cost of converting previous records to digital, learning to use the concept, thickness of sensor and lack of
59 universal use of digital radiography [4].

60 In recent years, a digital imaging system -Radiovisiography (RVG) has offered an alternative and instant
61 method for measurement of intraoral radiography [5]. It has been reported that RVG system provided
62 approximately an 80% reduction in radiation dosage in comparison with conventional X-ray films [6].

63 The present study was aimed at the comparative evaluation of conventional and digital radiography (RVG) to
64 measure alveolar bone loss in an attempt to help the clinician and practitioner to select the reliable radiographic
65 method for imaging and detection of alveolar bone loss.

66 2 (D)

67 II.

68 3 Methods

69 This study was carried out among fifty patients who were randomly selected from the OPD of Department of
70 Oral Medicine and Radiology in P.M.N.M Dental College and Hospital, Bagalkot. The study included 40 males
71 and 10 females aged between 20 -65 years.

72 4 Inclusion criteria

73 Patient having generalized mild to severe chronic periodontitis as assessed by measuring attachment loss and
74 categorized as mild: 1-2mm, moderate: 3-4 mm, severe: ? 5 mm.

75 5 Exclusion criteria

76 Patients with drifted teeth, supraerupted teeth and those who were contraindicated for any radiographic procedure
77 were excluded from the study.

78 A series of conventional bitewing radiographs (15,16, ???, ??5, ??6, ???, ??5, ??6, ???, ??5, ??6, ???) were
79 taken for each of fifty patients having chronic periodontitis using Satelec dental X-ray unit operated at 70 Kvp
80 and 8mA with a radiation exposure of 0.8 seconds. The film used was Kodak E -speed, number 2 of size 41x
81 31mm (Ekta speed, Eastman Kodak, Rochester, USA) and processing of the film was done manually using time
82 temperature method. Radiographs were mounted on xray viewer and alveolar bone loss was measured by keeping
83 divider on the CEJ to the most apical level of marginal bone. Later transparent ruler was used to evaluate
84 the distances between the two points of divider (Figure 1). Similarly a series of digital bitewing radiographs
85 (15,16, ???, ??5, ??6, ???, ??5, ??6, ???, ??5, ??6, ???) were taken for each of fifty patients by using RVG
86 of Kodak 5000 system and Satelec dental X-ray unit operated at 70 Kvp and 8mA with a radiation exposure
87 of 0.2 seconds. To ensure maximum hygiene, we covered the sensor with plastic sleeves and for each patient
88 a new plastic cover was used. The system we used in our study contained a charged coupled device or CCD
89 sensor. The measurement in RVG was done using Kodak 5000 digital software (Figure 2). Radiographically for
90 the measurement of bone levels from both methods we considered normal bone level less than 2mm from the
91 CEJ, and above that we measured as bone loss. Total 600 sites were measured and 3 readings were taken from
92 each site and mean of 3 readings is taken as a final readings. The Statistical software namely SPSS version 16.0
93 was used for data analysis. Paired t-test was performed on all the variables to evaluate between both the groups
94 at p ? 0.05. Relative agreement between the groups was done with Pearson's Correlation Coefficient.

6 III.

7 Results

Fifty subjects were included in the study and total number of sites measured was 600. The overall results show that the mean difference between both the conventional and digital bitewing radiographs was -0.4595 and the standard deviation difference were 0.2409 which is statistically significant (Table-1). It was observed that digital bitewing radiographs evaluated about 0.4mm greater bone loss on an average than conventional bitewing radiographs. By using paired t-test, results showed differences between right and left side sextants. In right maxilla, conventional images showed mean bone loss of 4.1mm, while digital images indicated a mean bone loss of 4.2mm, which was statistically significant. In the left maxilla conventional images showed a mean bone loss of 4.5mm while digital images showed an average loss of 4.6mm, which was statistically significant (Table-2).

Where as in right mandible conventional images showed averaged bone loss of 3.3mm, while digital images showed 4.0mm of averaged bone loss with a significant p-value. In left mandible conventional images showed averaged bone loss of 3.7mm, while digital images showed 4.0mm of averaged bone loss with a significant p-value (Table-3). RVG showed more bone loss in mandible than conventional radiographs but in maxilla both methods are showing almost similar measurement with a mean difference of 0.1mm only. Using Karl Pearson's correlation coefficient ($r = 0.9359$), there is a relative agreement of bone level measuring between conventional and digital bitewing radiographs (Table-4). Significant correlation was found between the RVG and conventional methods in both right and left sextant of both jaws ranging from $r = .82$ to $r = .91$ with a significant p-value (Table 5).

8 Discussion

Radiographs provide unique information about the status of the periodontium and a permanent record of the condition of the bone throughout the course of the disease. Radiographs aid the clinician in identifying the extent of destruction of alveolar bone, local contributing factors, and features of the periodontium that influence the prognosis [2].

The diagnosis of periodontal disease is primarily based on clinical examination. The clinical findings of periodontal osseous destruction can be confirmed by radiographic examination, but the radiographs on its own cannot help in diagnosing the disease [7].

Radiographic digital imaging systems like electronic probing system, subtraction radiography, CADIA (Computer Assisted Densitometric Image Analyses System), Dark field microscopy, DNA probes, Immunodiagnostic methods have been developed in recent times, which act as an adjunct in the precise diagnosis of periodontal disease [8]. Khocht et al (2003) stated that digital radiography offers many advantages over conventional methods [1]. It eliminates the need for film and film developing, and it allows for lower radiation exposure. The generated image is available immediately for evaluation on a computer screen and can be manipulated digitally to enhance viewing. In addition, digital tools are available to record electronic measurements and to cut, paste and colorize the image. The image can be easily filed on and retrieved from the hard disk or removable storage medium, or the images can be transferred electronically to third party carriers [9] Apart from these, one more advantage is the immediate observation of radiographic images. Only few digital radiography devices provide immediate viewing like charged coupled devices or CCDs [10]. However phosphorous plate technology requires placement of irradiated sensor in a processing device to scan it and put the information into a computer so that image can be viewed. In conventional radiographic techniques, the delay in reading the image usually forces the clinician to change his gloves and linger elsewhere as the radiographs undergo development [11].

RVG is also useful in educating and motivating the patient [12]. During implant placement, using conventional radiography is a major inconvenience, as the entire aseptic procedure is disrupted and time is wasted while the clinician awaits the development of the films several times during implant placement procedure [13]. It also allows the clinician to change contrast, enlarge images, place color enhancements or superimpose various textures on images [14].

However, as both advantages and disadvantages of any new invented device go hand in hand, the drawbacks include cost of the devices as well as converting previous records to digital, which are very high, thickness and rigidity of sensor that makes the patient uncomfortable, loss or breakage of sensor, which can prove very costly.

The clinical implications of radiography in the diagnosis of periodontal disease are twofold; to visualize the initial status of the bone tissue and to illustrate changes in bone tissue over time. When there are so many radiographic techniques, the clinician is in a dilemma as to which technique has to be used. This study was an attempt to help the clinician select the reliable radiographic method in imaging and detection of periodontal osseous destruction [15].

The results of the present study showed that overall the digital bitewing images averaged 0.4mm greater bone loss than did the conventional images with a significant p-value. Similarly, in a study conducted by Ahmed Kotch et al (2003) [1], digital radiography measured 0.3mm greater bone loss than conventional bitewings with significant p-value, which is relatively consistent with our results.

Given the overall difference between conventional and digital bitewing radiographs, we wanted to know if this difference was consistent across all sextants of the mouth. Therefore, we computed paired t-test for each of the four sextants available and our results showed measurement differences in RVG and conventional radiographs in all four sextants of the jaw.

9 CONCLUSION

156 RVG showed more bone loss in both maxilla and mandible than conventional radiographs, but in maxilla both
157 methods are showing almost similar measurement with a mean difference of 0.1mm only. However, in the study
158 of Ahmed Kotch et al (2003) digital radiographs observed bone loss only in the posterior mandibular region and
159 measurements of bone loss in the posterior maxillary region were similar between the two radiographic methods
160 [1].

161 Engebretson et al stated that there is no significant difference in conventional and digital radiographs as such
162 but digital radiographs are more accurate in measurements than film-based radiographs [16]. In our study also,
163 RVG showed more accuracy than conventional radiographs while measuring alveolar bone loss.

164 In the present study, we observed that in the normal clinical use, significant difference exists between alveolar
165 bone loss measurements on digital and conventional radiographs in several regions in the mouth. This difference
166 noted between the two imaging systems may be attributed to variations in measurements, which were done
167 manually in case of conventional radiographs and digitally in case of digital radiographs, because RVG was
168 showing 0.4mm greater bone loss than conventional radiographs while comparing the total samples. These
169 variations may be due to flexibility of the conventional radiograph film and sensor used in digital radiography.

170 V.

171 9 Conclusion

172 It can be hereby concluded that the digital radiographs have an upper hand when compared to conventional
173 radiographs in terms of alveolar bone loss. Although RVG has superior image recording capabilities compared to
conventional radiographs, its cost factor is an important point of consideration, which can limit its use. ^{1 2}



Figure 1: Figure 1 :

174

¹© 2015 Global Journals Inc. (US)

²Comparision between Conventional Radiography (IOPA) and Digital Radiography using BitewingTechnique in Detecting the Depth of Alveolar Bone Loss



2

Figure 2: Figure 2 :

1

Method	Mean	SD	Mean Diff	SD Diff	p-value
IOPA	3.9400	0.6666			
RVG	4.3995	0.6777	-0.4595	0.2409	0.000

Figure 3: Table 1 :

2

Method	Mean	SD	Mean Diff	SD Diff	p-value
IOPA (Right)	4.1067	0.7610			
RVG (Right)	4.2627	0.8200	-0.1560	0.4769	0.001
IOPA (Left)	4.5200	0.9102			
RVG (Left)	4.6933	0.8660	-0.1733	0.3784	0.000

Figure 4: Table 2 :

3

Method	Mean	SD	Mean Diff	SD Diff	p-value
IOPA (Right) RVG (Right)	3.3467 4.0587	0.7942 0.8505	-0.7120	0.4355	0.000
IOPA (Left) RVG (Left)	3.7867 4.5833	0.9803 0.9479	-0.7967	0.4869	0.000

Figure 5: Table 3 :

4

Methods	Conventional radiograph method (IOPA) r-value	p-value
Radiovisiography (RVG)	0.9359	0.0000

Figure 6: Table 4 :

5

Methods	Conventional radiograph method (IOPA) r-value	p-value
Radiovisiography (RVG) (Right maxilla)	0.8206	0.0000
Radiovisiography (RVG) (Left maxilla)	0.9104	0.0000
Radiovisiography (RVG) (Right mandible)	0.8620	0.0000
Radiovisiography (RVG) (Left mandible)	0.8730	0.0000

IV.

Figure 7: Table 5 :

-
- 175 [Dentomaxillofacial Radiology ()] , *Dentomaxillofacial Radiology* 2005. 34 p. .
- 176 [Vijay and Raghavan ()] , G Vijay , V Raghavan . *Radiology in Periodontics. Journal of Indian Academy of Oral*
177 *Medicine and Radiology* 2013. 25 (1) p. .
- 178 [Jenkins et al. ()] ‘A comparison of two radiographic assessment protocols for patients with periodontal disease’.
179 Wmm Jenkins , L M Brocklebank , S M Winning , A Donaldson , R M Strang . *Br Dent J* 2005. 198 (9) p. .
- 180 [Khocht et al. ()] ‘Comparison of direct digital and conventional intraoral radiographs in detecting alveolar bone
181 loss’. Ahmed Khocht , Maluinjanal , Kuang-Min Lynnharasty , Chang . *J Am Dent Assc* 2003. 134 p. .
- 182 [De et al. ()] ‘Detection of periodontal bone loss using cone beam CT and intraoral radiography’. K De , Faria
183 Vasconcelos , K M Evangelista , C D Rodrigues , C Estrela , To De Sousa , Mag Silva . *Dentomaxillofac*
184 *Radiol* 2012. 41 (1) p. .
- 185 [Esmaeli et al. ()] ‘Determination of vertical interproximal bone loss topography: correlation between indirect
186 digital radiographic measurement and clinical measurement’. F Esmaeli , A Shirmohammadi , M Faramarzie
187 , N Abolfazli , H Rasouli , S Fallahi . *Iran J Radiol* 2012. 9 p. .
- 188 [Taba et al. ()] ‘Diagnostic Biomarkers for Oral and Periodontal Diseases’. Mario Taba , Janet Kinney , Amy S
189 Kim , William V Giannobile . *Dent Clin North Am* 2005. 49 (3) p. 551.
- 190 [Bhambhani et al. ()] ‘Digitization and Its Futuristic Approach in Prosthodontics’. R Bhambhani , J Bhat-
191 tacharya , Saibal Kr Sen . *J Indian Prosthodont Soc* 2013. 13 (3) p. .
- 192 [Safi et al. ()] ‘Evaluation of alveolar crest bone loss via premolar bitewing radiographs: presentation of a new
193 method’. Y Safi , M Kadkhodazadeh , P Safai , M Esmaeelinejad , N Shamloo . *J Periodontal Implant Sci*
194 2014. 44 (5) p. .
- 195 [Engebretson et al. ()] ‘Gingival crevicular fluid levels of interleukin-1beta and glycemic control in patients with
196 chronic periodontitis and type 2 diabetes’. S P Engebretson , J Hey-Hadavi , F J Ehrhardt . *J Periodontol*
197 2004. 75 (9) p. .
- 198 [Ballrick et al. ()] ‘Image distortion and spatial resolution of a commercially available cone-beam computed
199 tomography machine’. J W Ballrick , J M Palomo , E Ruch , B D Amberman , M G Hans . *Am J Orthod*
200 *Dentofacial Orthop* 2008. 134 p. .
- 201 [Parissis et al.] N Parissis , A Kondylidou-Sidira , P Tsirlis , Patias . *Conventional radiographs vs digitized*
202 *radiographs: image quality assessment,*
- 203 [Morea et al. ()] ‘Quantitative analysis of bone density in direct digital radiographs evaluated by means of
204 computerized analysis of digital images’. C Morea , G C Dominguez , A Coutinho , I Chilvarquer .
205 *Dentomaxillofac Radiol* 2010. 39 (6) p. .
- 206 [Saad and Saad ()] ‘Radiation dose reduction during endodontic therapy: A new technique combining an apex
207 locator and digital imaging system (Radiovisiography)’. Yousef Saad , Ai-Nazhan Saad . *Journal of endodontics*
208 2000. 26 p. .
- 209 [Jeffcoat et al. (2000 1995)] ‘Radiographic diagnosis in periodontics’. M K Jeffcoat , W Chung , M S Reddy .
210 *Periodontology* 2000 1995. (7) p. .
- 211 [Wong et al. ()] ‘Radiographic study of ethnic variation in alveolar bone height among New Zealand dental
212 students’. Bkj Wong , J W Leichter , N P Chandler , M P Cullinan , D W Holborow . *J Periodontol* 2007. 78
213 p. .
- 214 [Chritensen ()] ‘Why switch to digital radiography’. G J Chritensen . *J Am Dent Assc* 2004. (135) p. .